


Certificate of Analysis and Release

Product: EpiScreen Plus

Batchnumber:		FP18EP04
Production date:	DD/MM/YYYY	8/08/2018
Expiry date:	DD/MM/YYYY	31/08/2020
Batch size:		100 kits

	<i>Result</i>	<i>Specification</i>
Reagent 1		
Osmolality (mOsm/kg):	234	210-260 mOsm/kg
pH:	6,85	6,70 - 6,90
Reagent 3		
Osmolality (mOsm/kg):	1066	930-1140 mOsm/kg
pH:	6,68	6,55 - 6,75
Reagent 4		
pH:	12,06	> 12
Reagent 6		
pH:	12,14	> 12
Functionality test:	OK	PASSED

Product released by Alexine Messiaen on 23/08/2018 (DD/MM/YYYY)



Alexine Messiaen