


Certificate of Analysis and Release

FertiCult Aspiration medium

| | | |
|------------------|------------|------------|
| Batchnumber: | | FP18FA19 |
| Production date: | DD/MM/YYYY | 07/09/2018 |
| Expiry date: | DD/MM/YYYY | 31/03/2020 |
| Total volume: | | 70000 ml |

| | <i>Results</i> | <i>Specifications</i> |
|--|----------------|-----------------------|
| pH: | 7,42 | 7.30-7.60 |
| Osmolality: | 277 | 270-290mOsm/kg |
| Bacterial screen: | PASSED | PASSED |
| Fungal screen: | PASSED | PASSED |
| Endotoxin: | <0,05 EU/ml | < 0,25EU/ml |
| Concentration heparin: | 2,265 | 2 - 3.75 g/liter |
| MEA (1-cell to expanded blastocyst within 96 hours): | 95,24% | ≥ 80% |

Product released by Alexine Messiaen on 24/09/2018 (DD/MM/YYYY)



Alexine Messiaen

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.