

Certificate of Analysis and Release

FertiCult IVF Medium

Batchnumber:	FP18FC10PRG	
Production date:	DD/MM/YYYY	07/08/2018
Expiry date:	DD/MM/YYYY	31/08/2019
Total volume:		20000 ml

	<i>Results</i>	<i>Specifications</i>
pH (@37°C and 5-6% CO ₂):	6,77	7.20-7.50
Osmolality:	279	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	<0,05 EU/ml	< 0,25EU/ml
Concentration HSA:	4,056	3.2 - 4.8 g/liter
Concentration gentamicin:	9,58	5 - 15 mg/liter
MEA (fertilization in FertiCult IVF medium, 1-cell to expanded blastocyst within 96 hours) :	86%	≥ 80%

Product released by Lenie Bocklandt on 29/08/2018 (DD/MM/YYYY)



Bocklandt lenie

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.