

Certificate of Analysis and Release

FertiCult Flushing Medium

Batchnumber:	FP18FL01PRG	
Production date:	DD/MM/YYYY	05/01/2018
Expiry date:	DD/MM/YYYY	31/07/2019
Total volume:		50000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,38	7.30-7.60
Osmolality:	279	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	<0,05 EU/ml	< 0,25EU/ml
Concentration HSA:	3,853	3.2 - 4.8 g/liter
Concentration gentamicin:	7,34	5 - 15 mg/liter
Mouse embryo test (1-cell):	90,48%	≥ 80%

Product released by Lenie Bocklandt on 22/01/2018 (DD/MM/YYYY)



Bocklandt lenie

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.