


Certificate of Analysis and Release

FertiCult Flushing Medium

Batchnumber:		FP18FL43PRG
Production date:	DD/MM/YYYY	17/09/2018
Expiry date:	DD/MM/YYYY	31/03/2020
Total volume:		40000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,45	7.30-7.60
Osmolality:	280	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	<0,05 EU/ml	< 0,25EU/ml
Concentration HSA:	4,046	3.2 - 4.8 g/liter
Concentration gentamicin:	7,53	5 - 15 mg/liter
MEA (1-cell to expanded blastocyst within 96 hours):	85,71%	≥ 80%

Product released by Alexine Messiaen on 02/10/2018 (DD/MM/YYYY)



Alexine Messiaen

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.