

Certificate of Analysis and Release

Product: FertiVit

Batchnumber :		FP18FVC05A
Production date :	DD/MM/YYYY	24/05/2018
Expiry date :	DD/MM/YYYY	31/05/2019
Total batchsize:		145 Kits

FertiVit C1	Result	Specification
pH :	7,35	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	19,243	15.6 -23.4 g/liter

FertiVit C2		
pH :	7,33	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	18,53	15.2 -22.8 g/liter

FertiVit C3		
pH :	7,32	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	18,015	14.4 -21.6 g/liter

FertiVit C4		
pH :	7,31	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	15,526	12.8 - 19.2 g/liter

FertiVit C5		
pH :	7,38	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	10,564	8.6 - 12.9 g/liter

FertiVit PI		
pH :	7,32	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml	< 0,25EU/mL
Osmolality:	283	270-290mOsm/kg
Concentration HSA:	19,864	16,0 - 24,0 g/liter

MEA (1-cell to expanded blastocyst within 96 hours):	100%	≥ 80%
--	------	-------

Product released by Lenie Bocklandt on 08/06/2018 (DD/MM/YYYY)



Bocklandt denie

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.