

## Certificate of Analysis and Release

### Product: FertiVit

Batchnumber :		FP18FVC09
Production date :	DD/MM/YYYY	29/08/2018
Expiry date :	DD/MM/YYYY	31/08/2019
Total batchsize:		149 Kits

<b>FertiVit C1</b>	<b>Result</b>	<b>Specification</b>
pH :	7,34	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	19,623	15.6 -23.4 g/liter

<b>FertiVit C2</b>		
pH :	7,38	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	19,245	15.2 -22.8 g/liter

<b>FertiVit C3</b>		
pH :	7,36	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	<0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	18,045	14.4 -21.6 g/liter


<b>FertiVit C4</b>		
pH :	7,37	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	16,058	12.8 - 19.2 g/liter

<b>FertiVit C5</b>		
pH :	7,38	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	10,152	8.6 - 12.9 g/liter

<b>FertiVit PI</b>		
pH :	7,35	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml	< 0,25EU/mL
Osmolality:	280	270-290mOsm/kg
Concentration HSA:	19,726	16,0 - 24,0 g/liter

MEA (1-cell to expanded blastocyst within 96 hours):	95,24%	≥ 80%
--	--------	-------

Product released by Alexine Messiaen on 13/09/2018 (DD/MM/YYYY)



Alexine Messiaen

**Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.**

*Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.*