

Certificate of Analysis and Release

Product: FertiVit

Batchnumber :		FP18FVW02
Production date :	DD/MM/YYYY	20/02/2018
Expiry date :	DD/MM/YYYY	28/02/2019
Total batchsize:		149 Kits

FertiVit W1	Result	Specification
pH :	7,25	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	0,1 EU/ml	< 0,25EU/mL
Osmolality:	1689	-
Concentration HSA:	17,361	13,9 - 20,9 g/liter
FertiVit W2		
pH :	7,26	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	0,1 EU/ml	< 0,25EU/mL
Osmolality :	1249	-
Concentration HSA:	17,178	14,4 - 21,6 g/liter
FertiVit W3		
pH :	7,27	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	0,1 EU/ml	< 0,25EU/mL
Osmolality :	835	805-850mOsm/kg
Concentration HSA:	17,618	14,9 - 22,3 g/liter
FertiVit W4		
pH :	7,27	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	0,1 EU/ml	< 0,25EU/mL
Osmolality :	555	535-565mOsm/kg
Concentration HSA:	19,54	15,4 - 23,1 g/liter
FertiVit W5		
pH :	7,27	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	0,1 EU/ml	< 0,25EU/mL
Osmolality :	419	405-435mOsm/kg
Concentration HSA:	19,922	15,7 - 23,6 g/liter
FertiVit W6		
pH :	7,23	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml	< 0,25EU/mL
Osmolality:	283	270-290mOsm/kg
Concentration HSA:	20,139	16,0 - 24,0 g/liter
Mouse Embryo Assay (blastocyst after 96h):	90,48%	≥ 80%

Product released by Justine Van Renterghem on 07/03/2018 (DD/MM/YYYY)



Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.