

Certificate of Analysis and Release

Product: SpermMar IgG

Batchnumber :		FP18G02
Production date :	DD/MM/YY	1/02/2018
Expiry date :	DD/MM/YY	31/08/2019
Number of kits in batch:		524 kits

Chemical analysis

pH:		8,21
Osmolality:		345
Concentration:		PASSED

Functional analysis

Positive control:		PASSED
Negative control:		PASSED

Product released by Justine Van Renterghem on 12/02/2018 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Justine Van Renterghem", written over a horizontal line.