

Certificate of Analysis and Release

Product: Hypo-osmotic swelling test

Batchnumber :		FP18HOS01
Production date :	DD/MM/YYYY	19/01/2018
Expiry date :	DD/MM/YYYY	31/01/2019
Batch size:		60 kits

	<i>Result</i>	<i>Specification</i>
Osmolality:	163 mOsm/kg	160-170mOsm/kg
pH:	7,29	7.2-7.6
Sterility (fungal and bacterial sterility test):	PASSED	PASSED

Product released by Lenie Bocklandt on 05/02/2018 (DD/MM/YYYY)



Bocklandt lenie