

## Certificate of Analysis and Release

### Hyaluronidase in FertiCult Flushing medium

Batchnumber:		FP18HY07
Production date:	DD/MM/YYYY	30/03/2018
Expiry date:	DD/MM/YYYY	31/03/2019
Total volume:		3300 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,51	7.30-7.60
Osmolality:	281	270-290mOsm/kg
Sterility (fungal and bacterial sterility test):	PASSED	PASSED
Endotoxin:	0,175 EU/ml	< 1 EU/ml
Concentration HSA:	4,219	3.2 - 4.8 g/liter
Enzyme activity:	95,71	40 - 160 IU/ml
Mouse embryo test (1-cell):	95,24%	≥ 80%

Product released by Lenie Bocklandt on 13/04/2018 (DD/MM/YYYY)



Bocklandt Lenie

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.