

## Certificate of Analysis and Release

### Hyaluronidase in FertiCult Flushing medium

|                  |            |            |
|------------------|------------|------------|
| Batchnumber:     |            | FP18HY08   |
| Production date: | DD/MM/YYYY | 12/04/2018 |
| Expiry date:     | DD/MM/YYYY | 30/04/2019 |
| Total volume:    |            | 3600 ml    |

|  | <i>Results</i> | <i>Specifications</i> |
|--|----------------|-----------------------|
| pH:  | 7,44           | 7.30-7.60             |
| Osmolality:                                      | 275            | 270-290mOsm/kg        |
| Sterility (fungal and bacterial sterility test): | PASSED         | PASSED                |
| Endotoxin:                                       | 0,101 EU/ml    | < 1 EU/ml             |
| Concentration HSA:                               | 4,013          | 3.2 - 4.8 g/liter     |
| Enzyme activity:                                 | 109            | 40 - 160 IU/ml        |
| Mouse embryo test (1-cell):                      | 100%           | ≥ 80%                 |

Product released by Justine Van Renterghem on 27/04/2018 (DD/MM/YYYY)



Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.