


Certificate of Analysis and Release

Hyaluronidase in FertiCult Flushing medium

Batchnumber:		FP18HY16
Production date:	DD/MM/YYYY	24/07/2018
Expiry date:	DD/MM/YYYY	31/07/2019
Total volume:		4500 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,43	7.30-7.60
Osmolality:	276	270-290mOsm/kg
Sterility (fungal and bacterial sterility test):	PASSED	PASSED
Endotoxin:	0,128 EU/ml	< 1 EU/ml
Concentration HSA:	4,281	3.2 - 4.8 g/liter
Enzyme activity:	109,47	40 - 160 IU/ml
MEA (1-cell to expanded blastocyst within 96 hours):	90,48%	≥ 80%

Product released by Alexine Messiaen on 08/08/2018 (DD/MM/YYYY)



Alexine Messiaen

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.