

Certificate of Analysis and Release

Hyaluronidase in FertiCult Flushing medium

Batchnumber:		FP18HY17
Production date:	DD/MM/YYYY	17/08/2018
Expiry date:	DD/MM/YYYY	31/08/2019
Total volume:		3600 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,53	7.30-7.60
Osmolality:	280	270-290mOsm/kg
Sterility (fungal and bacterial sterility test):	PASSED	PASSED
Endotoxin:	0,111 EU/ml	< 1 EU/ml
Concentration HSA:	3,894	3.2 - 4.8 g/liter
Enzyme activity:	135,92	40 - 160 IU/ml
MEA (1-cell to expanded blastocyst within 96 hours):	100%	≥ 80%

Product released by Lenie Bocklandt on 03/09/2018 (DD/MM/YYYY)



Bocklandt lenie

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.