

## Certificate of Analysis and Release

### 10% PVP in FertiCult Flushing medium

Batchnumber:		FP18PVP16
Production date:	DD/MM/YYYY	27/08/2018
Expiry date:	DD/MM/YYYY	31/05/2019
Total volume:		850 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,33	7.20-7.60
Osmolality:	300	280-320 mOsm/kg
Sterility (fungal and bacterial sterility test) :	PASSED	PASSED
Endotoxin:	0,0545 EU/ml	< 1 EU/ml
Concentration HSA:	3,29	3.2 - 4.8 g/liter
Viscosity (at 25°C)	200,4	> 170 cP
MEA (1-cell to expanded blastocyst within 96 hours):	90,48%	≥ 80%

Product released by Justine Van Renterghem on 11/09/2018 (DD/MM/YYYY)



*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*