

Certificate of Analysis and Release

Product: SpermFreeze

Batchnumber:		FP18SF03
Production date:	DD/MM/YYYY	1/02/2018
Expiry date:	DD/MM/YYYY	31/08/2019
Total number of kits:		167 kits

	Results	Specification
pH:	7,47	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (4 hours exposure):	98%	≥ 80%
Endotoxin:	<0,05 EU/ml	< 0,25 EU/mL
Concentration HSA:	4,128	3,2 - 4,8 g/liter

Product released by Justine Van Renterghem on 16/02/2018 (DD/MM/YYYY)



All blood products should be treated as potentially infectious. Source materials from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and HTLV I/II and non-reactive for HBsAg, HCV RNA, and HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents.