

Certificate of Analysis and Release

Product: SpermFreeze

| | | |
|-----------------------|------------|------------|
| Batchnumber: | | FP18SF07 |
| Production date: | DD/MM/YYYY | 23/04/2018 |
| Expiry date: | DD/MM/YYYY | 31/10/2019 |
| Total number of kits: | | 121 kits |

| | Results | Specification |
|---|--------------|-------------------|
| pH: | 7,52 | 7.20-7.60 |
| Bacterial screen: | PASSED | PASSED |
| Fungal screen: | PASSED | PASSED |
| Sperm survival Test (4 hours exposure): | 100% | ≥ 80% |
| Endotoxin: | < 0,05 EU/ml | < 0,25 EU/mL |
| Concentration HSA: | 3,919 | 3,2 - 4,8 g/liter |

Product released by Lenie Bocklandt on 08/05/2018 (DD/MM/YYYY)



Bocklandt denie

All blood products should be treated as potentially infectious. Source materials from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and HTLV I/II and non-reactive for HBsAg, HCV RNA, and HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents.