


Certificate of Analysis and Release

Product: SpermFreeze

Batchnumber:		FP18SF13
Production date:	DD/MM/YYYY	25/06/2018
Expiry date:	DD/MM/YYYY	31/12/2019
Total number of kits:		165 kits

	Results	Specification
pH:	7,49	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (4 hours exposure):	96%	≥ 80%
Endotoxin:	<0,05 EU/ml	< 0,25 EU/mL
Concentration HSA:	4,108	3,2 - 4,8 g/liter

Product released by Alexine Messiaen on 10/07/2018 (DD/MM/YYYY)



Alexine Messiaen

All blood products should be treated as potentially infectious. Source materials from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and HTLV I/II and non-reactive for HBsAg, HCV RNA, and HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents.