

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP18SIP02U
Production date :	DD/MM/YYYY	22/01/2018
Expiry date :	DD/MM/YYYY	31/07/2019
Total volume :		9400 ml

<i>Sil-select Plus: Upper layer</i>	<i>Result</i>	<i>Specification</i>
pH :	7,30	7.20 - 7.60
Osmolality :	314	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	0,170 EU/ml	< 0,5EU/mL
Sperm survival test (4 hours exposure)	100%	≥ 80%

Product released by Lenie Bocklandt on 06/02/2018 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Lenie Bocklandt".

Backlandt denie