

Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP18SIP04L
Production date :	DD/MM/YYYY	29/01/2018
Expiry date :	DD/MM/YYYY	31/07/2019
Total volume :		24750 ml

Sil-select Plus: Lower layer	<i>Result</i>	<i>Specification</i>
pH :	7,40	7.20 - 7.60
Osmolality :	336	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1063	1.1050 - 1.1150 g/mL
Viscosity in cP (25°C):	1,54	< 1.65 cP
Sperm survival test (4 hours exposure)	98%	≥ 80%
Sperm survival test (24 hours exposure)	93%	≥ 75%

Product released by Lenie Bocklandt on 13/02/2018 (DD/MM/YYYY)



Bocklandt denie