

## Certificate of Analysis and Release

### Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP18SIP09L
Production date :	DD/MM/YYYY	22/02/2018
Expiry date :	DD/MM/YYYY	31/08/2019
Total volume :		17350 ml

<b><i>Sil-select Plus: Lower layer</i></b>	<i>Result</i>	<i>Specification</i>
pH :	7,48	7.20 - 7.60
Osmolality :	333	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0.1 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1075	1.1050 - 1.1150 g/mL
Viscosity in cP (25°C):	1,53	< 1.65 cP
Sperm survival test (4 hours exposure)	99%	≥ 80%
Sperm survival test (24 hours exposure)	98%	≥ 75%

Product released by Justine Van Renterghem on 09/03/2018 (DD/MM/YYYY)

