

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

| | | |
|-------------------|------------|------------|
| Batchnumber : | | FP18SIP09U |
| Production date : | DD/MM/YYYY | 22/02/2018 |
| Expiry date : | DD/MM/YYYY | 31/08/2019 |
| Total volume : | | 15000 ml |

Sil-select Plus: Upper layer

| | <i>Result</i> | <i>Specification</i> |
|-----------------------------------------|---------------|----------------------|
| pH : | 7,42 | 7.20 - 7.60 |
| Osmolality : | 315 | 310-340mOsm/kg |
| Bacterial screen : | NEGATIVE | NEGATIVE |
| Fungal screen | NEGATIVE | NEGATIVE |
| Endotoxin : | <0,1 EU/ml | < 0,5EU/mL |
| Sperm survival test (4 hours exposure) | 100% | ≥ 80% |
| Sperm survival test (24 hours exposure) | 96% | ≥ 75% |

Product released by Justine Van Renterghem on 09/03/2018 (DD/MM/YYYY)

