

Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP18SIP10L
Production date :	DD/MM/YYYY	28/02/2018
Expiry date :	DD/MM/YYYY	31/08/2019
Total volume :		10000 ml

Sil-select Plus: Lower layer	<i>Result</i>	<i>Specification</i>
pH :	7,28	7.20 - 7.60
Osmolality :	333	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL
Density (g/ml):	1,1072	1.1050 - 1.1150 g/mL
Viscosity in cP (25°C):	1,50	< 1.65 cP
Sperm survival test (4 hours exposure)	100%	≥ 80%
Sperm survival test (24 hours exposure)	98%	≥ 75%

Product released by Justine Van Renterghem on 15/03/2018 (DD/MM/YYYY)

