

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP18SIP10U
Production date :	DD/MM/YYYY	28/02/2018
Expiry date :	DD/MM/YYYY	31/08/2019
Total volume :		10000 ml

<i>Sil-select Plus: Upper layer</i>	<i>Result</i>	<i>Specification</i>
pH :	7,27	7.20 - 7.60
Osmolality :	320	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL
Sperm survival test (4 hours exposure)	99%	≥ 80%
Sperm survival test (24 hours exposure)	99%	≥ 75%

Product released by Justine Van Renterghem on 15/03/2018 (DD/MM/YYYY)

