

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP18SIP11U
Production date :	DD/MM/YYYY	6/03/2018
Expiry date :	DD/MM/YYYY	30/09/2019
Total volume :		14000 ml

<i>Sil-select Plus: Upper layer</i>	<i>Result</i>	<i>Specification</i>
pH :	7,37	7.20 - 7.60
Osmolality :	321	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL
Sperm survival test (4 hours exposure)	100%	≥ 80%
Sperm survival test (24 hours exposure)	100%	≥ 75%

Product released by Justine Van Renterghem on 21/03/2018 (DD/MM/YYYY)

