

Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP18SIP13L
Production date :	DD/MM/YYYY	14/03/2018
Expiry date :	DD/MM/YYYY	30/09/2019
Total volume :		24625 ml

Sil-select Plus: Lower layer	<i>Result</i>	<i>Specification</i>
pH :	7,45	7.20 - 7.60
Osmolality :	319*	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1051	1.1050 - 1.1150 g/mL
Viscosity in cP (25°C):	1,53	< 1.65 cP
Sperm survival test (4 hours exposure)	96%	≥ 80%
Sperm survival test (24 hours exposure)	95%	≥ 75%

Product released by Lenie Bocklandt on 30/03/2018 (DD/MM/YYYY)



Bocklandt denie

* Based on the risk analysis performed, it has been decided that the minor deviation in osmolality is acceptable. Therefore, this batch was released under concession. More information on this matter is available upon request.

