

Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP18SIP21L
Production date :	DD/MM/YYYY	7/05/2018
Expiry date :	DD/MM/YYYY	30/11/2019
Total volume :		11100 ml

Sil-select Plus: Lower layer	<i>Result</i>	<i>Specification</i>
pH :	7,49	7.20 - 7.60
Osmolality :	333	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1094	1.1050 - 1.1150 g/mL
Viscosity in cP (25°C):	1,53	< 1.65 cP
Sperm survival test (4 hours exposure)	100%	≥ 80%
Sperm survival test (24 hours exposure)	98%	≥ 75%

Product released by Lenie Bocklandt on 22/05/2018 (DD/MM/YYYY)



Bocklandt denie