

Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP18SIP32L
Production date :	DD/MM/YYYY	20/08/2018
Expiry date :	DD/MM/YYYY	29/02/2020
Total volume :		20660 ml

Sil-select Plus: Lower layer	<i>Result</i>	<i>Specification</i>
pH :	7,33	7.20 - 7.60
Osmolality :	302	290-330mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	0,168 EU/ml	< 0,5EU/mL
Density (g/ml):	1,1091	1.1050 - 1.1150 g/mL
Viscosity in cP (25°C):	1,55	< 1.65 cP
Sperm survival test (4 hours exposure)	98%	≥ 80%
Sperm survival test (24 hours exposure)	95%	≥ 75%
Concentration HSA	0,426	0.3– 0.5 g/l

Product released by Lenie Bocklandt on 04/09/2018 (DD/MM/YYYY)



Bocklandt denie