


Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP18SIP35U
Production date :	DD/MM/YYYY	10/09/2018
Expiry date :	DD/MM/YYYY	31/03/2020
Total volume :		20955 ml

Sil-select Plus: Upper layer	<i>Result</i>	<i>Specification</i>
pH :	7,35	7.20 - 7.60
Osmolality :	292	280-310mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0.1 EU/mL	< 0,5EU/mL
Sperm survival test (4 hours exposure)	100%	≥ 80%
Sperm survival test (24 hours exposure)	93%	≥ 75%
Concentration HSA	2,328	1.8– 2.6 g/l

Product released by Alexine Messiaen on 25/09/2018 (DD/MM/YYYY)



Alexine Messiaen