

## Certificate of Analysis and Release

Product: Sil-Select Stock

Batchnumber :		FP18SIS16
Production date :	DD/MM/YYYY	15/06/2018
Expiry date :	DD/MM/YYYY	31/12/2019
Total volume :		60000 ml

	<i>Result</i>	<i>Specification</i>
pH :	7,26	7.20 - 7.60
Osmolality :	308	300 - 330 mOsm/kg
Sterility test bacteria:	NEGATIVE	NEGATIVE
Sterility test fungi:	NEGATIVE	NEGATIVE
Endotoxin:	< 0,1 EU/ml	< 0,5 EU/mL
Density (g/mL):	1,1215	1,1150-1,1250 g/mL
Viscosity in cP (25°C):	1,60	< 1,75 cP
Sperm survival test (4 hours exposure)	99%	≥ 80%
Sperm survival test (24 hours exposure)	100%	≥ 75%

Product released by Lenie Bocklandt on 02/07/2018 (DD/MM/YYYY)



Bocklandt lenie