

Certificate of Analysis and Release

Product: VitriFreeze ES media

Batchnumber :		FP18VFE09
Production date :	DD/MM/YYYY	7/06/2018
Expiry date :	DD/MM/YYYY	30/06/2019
Total batchsize:		200 Kits

Pre-incubation medium	Result	Specification
pH :	7,33	7.2 - 7.4
Osmolality :	280	270-295mOsm/kg
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1EU/ml	< 0,25EU/mL
Concentration HSA:	19,783	16,0 - 24,0 g/liter

VitriFreeze ES medium 1	Result	Specification
pH :	7,25	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	17,337	14,4 - 21,6 g/liter

VitriFreeze ES medium 2	Result	Specification
pH :	7,22	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA :	15,947	12,8 - 19,2 g/liter

VitriFreeze ES medium 3	Result	Specification
pH :	7,31	7.2-7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Concentration HSA :	10,005	8,3 - 12,4 g/liter

MEA (1-cell to expanded blastocyst within 96 hours):	90%	≥ 80%
--	-----	-------

Product released by Lenie Bocklandt on 22/06/2018 (DD/MM/YYYY)



Bocklandt lenie

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.