

Certificate of Analysis and Release

Product: VitriThaw ES media

Batchnumber :		FP18VTE03
Production date :	DD/MM/YYYY	31/01/2018
Expiry date :	DD/MM/YYYY	31/01/2019
Total batchsize:		199 Kits

VitriThaw ES medium 5	Result	Specification
pH :	7,22	7.2 - 7.4
Osmolality :	281	270-295mOsm/kg
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	22,176	16,0 - 24,0 g/liter

VitriThaw ES medium 4		
pH :	7,24	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Osmolality :	418	405-435mOsm/kg
Concentration HSA:	19,747	15.5 - 23.3 g/liter

VitriThaw ES medium 3		
pH :	7,24	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Osmolality :	549	535-565mOsm/kg
Concentration HSA :	20,917	15.2 - 22.8 g/liter

VitriThaw ES medium 2		
pH :	7,21	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Osmolality :	826	805-850mOsm/kg
Concentration HSA :	19,416	14.5 - 21.7 g/liter

VitriThaw ES medium 1		
pH :	7,28	7.2-7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Concentration HSA :	17,997	13.2 - 19.8 g/liter
Mouse Embryo Assay (blastocysts after 96h)	100%	≥ 80%

Product released by Justine Van Renterghem on 15/02/2018 (DD/MM/YYYY)

