

Certificate of Analysis and Release

Product: VitriThaw ES media

Batchnumber :		FP18VTE06
Production date :	DD/MM/YYYY	16/04/2018
Expiry date :	DD/MM/YYYY	30/04/2019
Total batchsize:		199 Kits

VitriThaw ES medium 5	Result	Specification
pH :	7,25	7.2 - 7.4
Osmolality :	279	270-295mOsm/kg
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,05 EU/ml	< 0,25EU/mL
Concentration HSA:	20,621	16,0 - 24,0 g/liter

VitriThaw ES medium 4		
pH :	7,32	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,05 EU/ml	< 0,25EU/mL
Osmolality :	416	405-435mOsm/kg
Concentration HSA:	20,205	15.5 - 23.3 g/liter

VitriThaw ES medium 3		
pH :	7,36	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1EU/ml	< 0,25EU/mL
Osmolality :	543	535-565mOsm/kg
Concentration HSA :	19,637	15.2 - 22.8 g/liter

VitriThaw ES medium 2		
pH :	7,33	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Osmolality :	832	805-850mOsm/kg
Concentration HSA :	18,25	14.5 - 21.7 g/liter

VitriThaw ES medium 1		
pH :	7,34	7.2-7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA :	16,588	13.2 - 19.8 g/liter
Mouse Embryo Assay (blastocysts after 96h)	90,48	≥ 80%

Product released by Justine Van Renterghem on 02/05/2018 (DD/MM/YYYY)



Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.