

Certificate of Analysis and Release

Product: VitriThaw ES media

Batchnumber :		FP18VTE10
Production date :	DD/MM/YYYY	7/06/2018
Expiry date :	DD/MM/YYYY	30/06/2019
Total batchsize:		200 Kits

VitriThaw ES medium 5	Result	Specification
pH :	7,33	7.2 - 7.4
Osmolality :	280	270-295mOsm/kg
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	19,783	16,0 - 24,0 g/liter
VitriThaw ES medium 4		
pH :	7,28	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Osmolality :	414	405-435mOsm/kg
Concentration HSA:	17,916	15.5 - 23.3 g/liter
VitriThaw ES medium 3		
pH :	7,29	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Osmolality :	551	535-565mOsm/kg
Concentration HSA :	17,693	15.2 - 22.8 g/liter
VitriThaw ES medium 2		
pH :	7,27	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Osmolality :	831	805-850mOsm/kg
Concentration HSA :	16,245	14.5 - 21.7 g/liter
VitriThaw ES medium 1		
pH :	7,30	7.2-7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Concentration HSA :	15,268	13.2 - 19.8 g/liter
MEA (1-cell to expanded blastocyst within 96 hours):	90%	≥ 80%

Product released by Lenie Bocklandt on 22/06/2018 (DD/MM/YYYY)



Bocklandt lenie

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.