

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:
Embryo Transfer Cathether Emtrac A #4219 A

LOT number: 20180110

STERILIZATION date: 20-01-2018

EXPIRY date: 10-01-2022

STERILIZATION technique: Gamma Irradiation

STERILIZATION dose: 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:
24-01-2018