

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code:

Embryo Transfer Cathether Emtrac A #4219 A

LOT number: 20141001**STERILIZATION date:** 05-10-2014**EXPIRY date:** 2018-10-01**STERILIZATION technique:** gamma irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



K. Daemen
Gynetics Medical Products

SIGNATURE date:
22-10-2014