

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

ITEM name and code: #

Embryo Transfer Catheter Emtrac Set #4219 Set

LOT number: 20160302

STERILIZATION date: 06-03-2016

EXPIRY date: 2020-03-02

STERILIZATION technique: gamma irradiation

STERILIZATION dose: 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



K. Daemen  
Gynetics Medical Products

SIGNATURE date:  
21-03-2016