

# Gynetics Medical Products CERTIFICATE OF QUALITY

**ITEM name and code:**

Intra-Uterine Insemination Cannula with memory #4220 IUI Memo

**LOT number:** 20180213**STERILIZATION date:** 17-02-2018**EXPIRY date:** 13-02-2022**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

## QUALITY CONTROL TESTING PROCEDURES

| TEST                  | SPECIFICATION  | CONCLUSION |
|-----------------------|----------------|------------|
| Endotoxin Assay (LAL) | <0.5 EU/device | Pass       |

| TEST               | SPECIFICATIONS | CONCLUSION |
|--------------------|----------------|------------|
| Mouse Embryo Assay | >= 80%         | Pass       |
| MEA Control        | >= 80%         | Pass       |

Greetje Hermans  
Gynetics Medical Products**SIGNATURE date:**  
07-03-2018