

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**

Intra-Uterine Insemination Cannula with memory #4220 IUI Memo

**LOT number:** 20180214**STERILIZATION date:** 17-02-2018**EXPIRY date:** 14-02-2022**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass

Greetje Hermans  
Gynetics Medical Products**SIGNATURE date:**  
07-03-2018