

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code:

Intra-Uterine Insemination Cannula with memory #4220 IUI Memo

LOT number: 20150107**STERILIZATION date:** 11-01-2015**EXPIRY date:** 2019-01-07**STERILIZATION technique:** gamma irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



K. Daemen
Gynetics Medical Products

SIGNATURE date:
28-01-2015