

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code: #

Intra-Uterine Insemination Cannula with memory #4220 IUI Memo

LOT number: 20160302**STERILIZATION date:** 06-03-2016**EXPIRY date:** 2020-03-02**STERILIZATION technique:** gamma irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass

K. Daemen
Gynetics Medical Products**SIGNATURE date:**
21-03-2016