

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code:
Intra-Uterine Insemination Cannula #4220 IUI

LOT number: 20141203

STERILIZATION date: 8-12-2014

EXPIRY date: 2018-12-03

STERILIZATION technique: Gamma irradiation

STERILIZATION dose: 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

| TEST | SPECIFICATION | CONCLUSION |
|-----------------------|----------------|------------|
| Endotoxin Assay (LAL) | <0.5 EU/device | Pass |

| TEST | SPECIFICATIONS | CONCLUSION |
|--------------------|----------------|------------|
| Mouse Embryo Assay | >= 80% | Pass |
| MEA Control | >= 80% | Pass |



K. Daemen
Gynetics Medical Products

SIGNATURE date:
07-01-2015