

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code:
Intra-Uterine Insemination Cannula #4220 IUI

LOT number: 20150112

STERILIZATION date: 18-01-2015

EXPIRY date: 2019-01-12

STERILIZATION technique: gamma irradiation

STERILIZATION dose: 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass

K. Daemen
Gynetics Medical Products

SIGNATURE date:
11-02-2015