

# Gynetics Medical Products CERTIFICATE OF QUALITY

**ITEM name and code:**

Intra-Uterine Insemination Cannula #4220 IUI

**LOT number:** 20150210**STERILIZATION date:** 15-02-2015**EXPIRY date:** 2019-02-10**STERILIZATION technique:** gamma irradiation**STERILIZATION dose:** 25 - 39 kGray

## QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass

K. Daemen  
Gynetics Medical Products**SIGNATURE date:**  
04-03-2015