

# Gynetics Medical Products CERTIFICATE OF QUALITY

**ITEM name and code:**  
Denuding Pipette #3000 Strip-ID125/10

**LOT number:** 20160227

**STERILIZATION date:** 06-03-2016

**EXPIRY date:** 2020-02-27

**STERILIZATION technique:** gamma irradiation

**STERILIZATION dose:** 25 - 39 kGray

## QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass

K. Daemen  
Gynetics Medical Products

**SIGNATURE date:**  
21-03-2016