

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code: *
Denuding Pipette #3000 Strip-ID135/10

LOT number: 20160419

STERILIZATION date: 24-04-2016

EXPIRY date: 2020-04-19

STERILIZATION technique: gamma irradiation

STERILIZATION dose: 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass

K. Daemen
Gynetics Medical Products

SIGNATURE date:
09-05-2016