

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:** \*

Denuding Pipette #3000 Strip-ID170/10

**LOT number:** 20150923

**STERILIZATION date:** 27-09-2015

**EXPIRY date:** 2019-09-23

**STERILIZATION technique:** gamma irradiation

**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

| TEST                  | SPECIFICATION  | CONCLUSION |
|-----------------------|----------------|------------|
| Endotoxin Assay (LAL) | <0.5 EU/device | Pass       |

| TEST               | SPECIFICATIONS | CONCLUSION |
|--------------------|----------------|------------|
| Mouse Embryo Assay | >= 80%         | Pass       |
| MEA Control        | >= 80%         | Pass       |



K. Daemen  
Gynetics Medical Products

**SIGNATURE date:**  
12-10-2015