

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code:
Denuding Pipette #3000 Strip-ID175/10

LOT number: 20160328

STERILIZATION date: 27-03-2016

EXPIRY date: 2020-03-28

STERILIZATION technique: gamma irradiation

STERILIZATION dose: 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



K. Daemen
Gynetics Medical Products

SIGNATURE date:
13-04-2016