

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**  
Denuding Pipette #3000 Strip-ID275/10

**LOT number:** 20151021

**STERILIZATION date:** 25-10-2015

**EXPIRY date:** 2019-10-21

**STERILIZATION technique:** gamma irradiation

**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



K. Daemen  
Gynetics Medical Products

**SIGNATURE date:**  
09-11-2015