

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:
Denuding Pipette #3000 Strip-ID135/10

LOT number: 20151125

STERILIZATION date: 29-11-2015

EXPIRY date: 2019-11-25

STERILIZATION technique: gamma irradiation

STERILIZATION dose: 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



K. Daemen
Gynetics Medical Products

SIGNATURE date:
11-12-2015